



Ref. Certif. No.

DE 3 - 40791

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

### CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product	General Medical Devices Vaporizer
Name and address of the applicant	STORZ & BICKEL GMBH & CO. KG In Grubenäcker 5-9 78532 Tuttlingen, GERMANY
Name and address of the manufacturer	STORZ & BICKEL GMBH & CO. KG In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY
Name and address of the factory	STORZ & BICKEL GMBH & CO. KG In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY
Ratings and principal characteristics	Rated Input Voltage: 12 DC Rated Input Power: 36 W Protection Class: internally powered Degree of protection against electric shock: BF Degree of waterprotection: IP22
Trade mark (if any)	vapormed
Model/type Ref.	Mighty Medic
Additional information (if necessary)	Certificate DE 3 - 40357 issued 2015-07-20 is replaced by this version due to technical changes.
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-11:2015
as shown in the Test Report Ref. No. which forms part of this certificate	713112373

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2017-07-19  
CB 17 07 63179 063

Stefan Reis



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service